



Celsius Refrigerator Temperature Log

DAYS 1-15

Month/Year _____ VFC PIN _____

Facility Name _____

Storage Unit _____ Reviewer Signature _____

Monitor temperatures closely!

1. Write staff initials in "Staff Initials," and note the time in "Exact Time."
2. Record temperatures twice each workday.
3. Record the min/max temperatures once each workday – preferably in the morning.
4. Write the temperature in the row that corresponds to the refrigerator's temperature.
5. For out-of-range temperatures, see instructions to the right.
6. Bimonthly, download and print data logger data.
7. Maintain printed data logger data and temperature logs together for a minimum of three years.

Take action if temperatures are out of range - too warm (above 8°C) or too cold (below 2°C).

1. Label exposed vaccine "do not use," and store it under proper conditions as quickly as possible.
Do not discard vaccines unless directed to by the Iowa Immunization Program and/or the manufacturer(s).
2. Record the out-of-range temperatures and the room temperature in the "Action" area on the bottom of the log.
3. Notify the clinic vaccine coordinator and call the Iowa Immunization Program at 1-800-831-6293.
4. Document the action taken on the Emergency Response Worksheet and Troubleshooting Record.

Day of Month	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	
Staff Initials																
Exact Time	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM
Alarm (Yes/No)																
Min/Max Temps (since previous reading)																
Danger! Temperatures above 8°C are too warm! Write out-of-range temperatures and room temperature on the lines below and call the Immunization Program.																
TEMPERATURES	8°C															
	7°C															
	6°C															
	Aim for 5°C → 5°C															
ACCEPTABLE	4°C															
	3°C															
	2°C															
	Danger! Temperatures below 2°C are too cold! Write out-of-range temperatures and room temperature on the lines below and call the Immunization Program.															
ACTION	Write any out-of-range temps (above 8°C or below 2°C) here															
	Room Temperature															

DOWNLOAD DATA LOGGER DATA



Celsius Refrigerator Temperature Log

DAYS 16-31

Month/Year _____ VFC PIN _____

Facility Name _____

Storage Unit _____ Reviewer Signature _____

Monitor temperatures closely!

1. Write staff initials in "Staff Initials," and note the time in "Exact Time."
2. Record temperatures twice each workday.
3. Record the min/max temperatures once each workday – preferably in the morning.
4. Write the temperature in the row that corresponds to the refrigerator's temperature.
5. For out-of-range temperatures, see instructions to the right.
6. Bimonthly, download and print data logger data.
7. Maintain printed data logger data and temperature logs together for a minimum of three years.

Take action if temperatures are out of range - too warm (above 8°C) or too cold (below 2°C).

1. Label exposed vaccine "do not use," and store it under proper conditions as quickly as possible.
Do not discard vaccines unless directed to by the Iowa Immunization Program and/or the manufacturer(s).
2. Record the out-of-range temperatures and the room temperature in the "Action" area on the bottom of the log.
3. Notify the clinic vaccine coordinator and call the Iowa Immunization Program at 1-800-831-6293.
4. Document the action taken on the Emergency Response Worksheet and Troubleshooting Record.

Day of Month	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Staff Initials																
Exact Time	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM
Alarm (Yes/No)																
Min/Max Temps (since previous reading)																
Danger! Temperatures above 8°C are too warm! Write out-of-range temperatures and room temperature on the lines below and call the Immunization Program.																
TEMPERATURES	8°C															
	7°C															
	6°C															
	Aim for 5°C															
ACCEPTABLE	4°C															
	3°C															
	2°C															
Danger! Temperatures below 2°C are too cold! Write out-of-range temperatures and room temperature on the lines below and call the Immunization Program.																
ACTION	Write any out-of-range temps (above 8°C or below 2°C) here															
	Room Temperature															

DOWNLOAD DATA LOGGER DATA